



Attestation of Conformity

No. 190207-01

This attestation of conformity corresponds to the standard
 DIN EN ISO/IEC 17050-1
 „Conformity assessment – Supplier’s declaration of conformity- Part 1: General requirements”.

We, the manufacturer: EPHY-MESS GmbH
 Berta-Cramer-Ring 1
 65205 Wiesbaden
 Germany

declare under sole responsibility that the product (component)

model: Temperature sensor
 Type: PR-SPA-EX-NWT
 II 2G Ex eb IIC Gb
 II 2D Ex tb IIIC Db
 II 2G Ex ia IIC Gb
 II 2D Ex ia IIIC Db

to which this declaration relates is in conformity with the following standard(s) or other normative documents:

- EN 60079-0:2012+A11:2013
- EN 60079-7:2015
- EN 60079-11:2012
- EN 60079-31:2014
- IBExU 14 ATEX 1281 U, 1. Issue 01.02.2019

in accordance with the provisions of the directive(s):

- Directive 2014/34/EU (ATEX equipment directive)
- Directive 2011/65/EU (RoHS II)

EPHY MESS
 Gesellschaft für Elektro-Phys. Meßgeräte mbH
 Berta - Cramer - Ring 1
 65205 Wiesbaden-Delkenheim
 Telefon (06122) 9 22 8-0
 Telefax (06122) 9 22 8-99



Wiesbaden, 07.02.2019
 city / date

Bernd Horneffer
 Bernd Horneffer, ATEX-Officer

L. Sennerich
 L. Sennerich, ATEX-Officer

Commercial registries
 Registration court Wiesbaden
 HRB-Nr.: 22 729
 VAT REG No. DE113842153
 founded:1955
 General manager
 Dipl.-Kfm. Andreas Becker

Registered office
 Berta-Cramer-Ring 1
 65205 Wiesbaden, Germany
 Phone: +49 61 22 92 28 -0
 Fax: +49 61 22 92 28 -99
 E-Mail: info@ephy-mess.de

Quality assurance system
 Certified according to
 DIN ISO 9001
 DIN ISO 14001

